





Prüfbericht-Nr.: <i>Test Report No.:</i>	50343522 001	Auftrags-Nr.: <i>Order No.:</i>	244161359	Seite 1 von 126 <i>Page 1 of 126</i>	
Kunden-Referenz-Nr.: <i>Client Reference No.:</i>	405703	Auftragsdatum: <i>Order date:</i>	2019-07-29		
Auftraggeber: <i>Client:</i>	Qingdao Haier Biomedical Co., Ltd. Haier Industrial Park, Economic Technology Development Zone, Qingdao 266510, P.R. China				
Prüfgegenstand: <i>Test item:</i>	Pharmaceutical Refrigerator				
Bezeichnung / Typ-Nr.: <i>Identification / Type No.:</i>	HYC-68A, HYC-68				
Auftrags-Inhalt: <i>Order content:</i>	Issue test report				
Prüfgrundlage: <i>Test specification:</i>	EN 60335-1:2012 + A11+ A13 EN 60335-2-89:2010 + A1+ A2 EN 62233:2008				
Wareneingangsdatum: <i>Date of receipt:</i>	2019-10-19				
Prüfmuster-Nr.: <i>Test sample No.:</i>	244161359-001				
Prüfzeitraum: <i>Testing period:</i>	2019-10-20 to 2019-11-13				
Ort der Prüfung: <i>Place of testing:</i>	Qingdao Profound Testing Technology Service Co., Ltd.				
Prüflaboratorium: <i>Testing laboratory:</i>	TÜV Rheinland (Shanghai) Co., Ltd.				
Prüfergebnis*: <i>Test result*:</i>	Pass				
geprüft von / tested by:	kontrolliert von / reviewed by:				
2020-02-26 Allen Fang / PE 	2020-02-26 Nick Chen / Review 				
2020-02-26 Lucy Qi / Trainee 					
Datum <i>Date</i>	Name / Stellung <i>Name / Position</i>	Unterschrift <i>Signature</i>	Datum <i>Date</i>	Name / Stellung <i>Name / Position</i>	Unterschrift <i>Signature</i>
Sonstiges / Other: This report contains IEC report (page 2 to 112) and EN deviations (page 113 to 126). Attachment 1: Test equipment list (3 pages) Attachment 2: Photo documentation (19 pages)					
Zustand des Prüfgegenstandes bei Anlieferung: <i>Condition of the test item at delivery:</i>		Prüfmuster vollständig und unbeschädigt <i>Test item complete and undamaged</i>			
* Legende: 1 = sehr gut 2 = gut 3 = befriedigend 4 = ausreichend 5 = mangelhaft P(ass) = entspricht o.g. Prüfgrundlage(n) F(ail) = entspricht nicht o.g. Prüfgrundlage(n) N/A = nicht anwendbar N/T = nicht getestet Legend: 1 = very good 2 = good 3 = satisfactory 4 = sufficient 5 = poor P(ass) = passed a.m. test specification(s) F(ail) = failed a.m. test specification(s) N/A = not applicable N/T = not tested					
Dieser Prüfbericht bezieht sich nur auf das o.g. Prüfmuster und darf ohne Genehmigung der Prüfstelle nicht auszugsweise vervielfältigt werden. Dieser Bericht berechtigt nicht zur Verwendung eines Prüfzeichens. <i>This test report only relates to the a. m. test sample. Without permission of the test center this test report is not permitted to be duplicated in extracts. This test report does not entitle to carry any test mark.</i>					
V04					

Test Report issued under the responsibility of:



TEST REPORT IEC 60335-2-89 Safety of household and similar electrical appliances Part 2: Particular requirements for commercial refrigerating appliances with an incorporated or remote refrigerant condensing unit or compressor	
Report Number	50343522 001
Date of issue	See cover page
Total number of pages	See cover page
Name of Testing Laboratory preparing the Report	TÜV Rheinland (Shanghai) Co., Ltd. No.177, 178, Lane 777 West Guangzhong Road, Jing'an District, Shanghai China
Applicant's name	Qingdao Haier Biomedical Co., Ltd.
Address	Haier Industrial Park, Economic Technology Development Zone, Qingdao 266510, P.R. China
Test specification:	
Standard	IEC 60335-2-89 :2010, AMD1:2012, AMD2:2015 in conjunction with IEC60335-1:2010, AMD1:2013, AMD2:2016
Test procedure	Issue test report
Non-standard test method	N/A
Test Report Form No.	IEC60335_2_89I
Test Report Form(s) Originator	IMQ S.p.A.
Master TRF	Dated 2017-09
Copyright © 2017 IEC System of Conformity Assessment Schemes for Electrotechnical Equipment and Components (IECEE System). All rights reserved. This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context. This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.	
General disclaimer: The test results presented in this report relate only to the object tested.	