

TEST REPORT
IEC 61010-1
Safety requirements for electrical equipment for measurement,
control, and laboratory use
Part 1: General requirements

Report Number : 64.111.20.00608.01 Rev.00

Date of issue : 2020-03-25

Total number of pages..... : 89

Name of Testing Laboratory preparing the Report..... : TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch

Applicant's name : Qingdao Haier Biomedical Co., Ltd.

Address : Haier Industrial Park, Economic Technology Development Zone, Qingdao, 266510, P.R. China

Test specification:

Standard..... : IEC 61010-1:2010, AMD1:2016

Test procedure : CE-LVD

Non-standard test method : N/A

Test Report Form No. : IEC61010_1L

Test Report Form(s) Originator : VDE Testing and Certification Institute

Master TRF..... : 2017-10-20

Copyright © 2017 IEC System of Conformity Assessment Schemes for Electrotechnical Equipment and Components (IECEE System). All rights reserved.

This publication may be reproduced in whole or in part for non-commercial purposes as long as the IECEE is acknowledged as copyright owner and source of the material. IECEE takes no responsibility for and will not assume liability for damages resulting from the reader's interpretation of the reproduced material due to its placement and context.

If this Test Report Form is used by non-IECEE members, the IECEE/IEC logo and the reference to the CB Scheme procedure shall be removed.

This report is not valid as a CB Test Report unless signed by an approved CB Testing Laboratory and appended to a CB Test Certificate issued by an NCB in accordance with IECEE 02.

General disclaimer:

The test results presented in this report relate only to the object tested.

This report shall not be reproduced, except in full, without the written approval of the Issuing CB Testing Laboratory. The authenticity of this Test Report and its contents can be verified by contacting the NCB, responsible for this Test Report.

Test item description :	Pharmaceutical Refrigerator	
Trade Mark :	Haier	
Manufacturer	Same as applicant	
Model/Type reference :	HY-118, HY-118A	
Ratings :	220-240V~; 50/60Hz; R600a, Class I, Pressure of high/low side: 12,5 bar/0,8bar, 230W/1,5A for HYC-118A, 190W/1,4A for HYC-118	
Responsible Testing Laboratory (as applicable), testing procedure and testing location(s):		
<input checked="" type="checkbox"/>	Testing Laboratory:	
Testing location/ address	TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch Location 1: 5F, Communication Building, 163 Pingyun Rd, Huangpu Ave. West Guangzhou 510656 P. R. China Location 2: B1 Floor, No. 3 Chuangqi Building, TusPark, 63 Chuangqi Road, Shilou Town, Panyu District, Guangzhou 511447, P.R. China	
Tested by (name, function, signature) :	Jason Kong, Project Handler	<i>Jason kong</i>
Approved by (name, function, signature) .. :	Wang Xujun, Designated Reviewer	<i>Wang Xujun</i>
<input type="checkbox"/>	Testing procedure: CTF Stage 1:	N/A
Testing location/ address	N/A	
Tested by (name, function, signature) :	N/A	
Approved by (name, function, signature) .. :	N/A	
<input type="checkbox"/>	Testing procedure: CTF Stage 2:	N/A
Testing location/ address	N/A	
Tested by (name + signature) :	N/A	
Witnessed by (name, function, signature) . :	N/A	
Approved by (name, function, signature) .. :	N/A	
<input type="checkbox"/>	Testing procedure: CTF Stage 3:	N/A
<input type="checkbox"/>	Testing procedure: CTF Stage 4:	N/A
Testing location/ address	N/A	
Tested by (name, function, signature) :	N/A	
Witnessed by (name, function, signature) . :	N/A	
Approved by (name, function, signature) .. :	N/A	
Supervised by (name, function, signature) :	N/A	



List of Attachments (including a total number of pages in each attachment)		
Document No.	Documents included / attached to this report (description)	Page No.
Attachment No.1	Test report of EN 61010-2-011:2017	25

Documents referenced by this report (available on request):		
Document Name or No.	Documents description	Page No.
--	--	--